



बी.पी. कोइराला मेमोरियल क्यान्सर अस्पताल

भरतपुर, चितवन, नेपाल

पो.बक्स नं. ३४, फोन नं. ०५६-५२४५०१, फ्याक्स नं. ०५६-५२३०४७, वेबसाइट: www.bpkmch.org.np




प.सं.

चलानी नं.

दररेट पेश गर्ने सम्बन्धि सूचना ।

प्रकाशित मिति : २०७९/१/१५

यस अस्पतालको Department of Radiation Oncology का लागि SRS/SRBT का लागि आवश्यक Base Plate Set बोलपत्र माध्यमबाट खरिद गर्न लागत अनुमान तयार पार्ने प्रयोजनका लागि ईच्छुक, योग्य कम्पनी/फर्म तथा सप्लायर्सहरुले आवश्यक कागजात सहित यसैसाथ संलग्न प्राविधिक स्पेसिफिकेशन बमोजिमको आर्थिक प्रस्ताव (७सात) दिनभित्र कार्यालय समयभित्र अस्पताल प्रशासन शाखामा सोभै वा Email ID directorbpmch@gmail.com or bpmchprocurement@gmail.com मा पठाउनुहुन अनुरोध छ । बिस्तृत जानकारीका लागि अस्पतालको खरिद उप-शाखामा सम्पर्क गर्नुहुन सूचित गरिन्छ ।


डा. प. ज. कुमार गौतम
कार्यकारी निर्देशक



B. P. Koirala Memorial Cancer Hospital

Chitwan

Fiscal Year 2078/079

Supply, Delivery of Immobilization equipments and accessories

A	<p>A complete Stereotactic Radiosurgery (SRS) system is used to immobilize the patients securely, precisely, and reproducibly for the safe and accurate delivery of high precision SRS treatments. The system must be USFDA and CE-approved.</p>	1 set	
	<p>Technical Specification: A complete set of SRS systems with its standard accessories.</p>		
	<p>Should provide SRS base plate, standard SRS thermoplastic head masks, open eye and mouth (4nos), head and shoulder (2nos) mask, shim system with changeable mask fixing frames, Bite Locks, adjustable hand grips, appropriate headrest and cushion (if required) as.</p>		
	<p>User Training : Must provide user training (including how to use and maintain the equipment's) to the operator</p>		
	<p>Warranty The warranty period shall be 1 year</p>		
B	<p>A complete Stereotactic Body Radiation Therapy (SBRT) system: is used to immobilize the patients securely, precisely, and reproducibly for the safe and accurate delivery of high precision SBRT treatments for tumors outside the CNS that require either a single high dose radiation treatment or a few fractionated radiation treatments. The system must be USFDA and CE approved.</p>	1 set	
	<p>Technical Specifications: A complete set of SBRT systems with its standard accessories</p>		
	<p>Accessories:</p>		
	<p>Must provide vacuum bags, cushions, elevation blocks, foot holders, knee and abdominal bridges, standard thermoplastic abdomen, thorax, pelvic masks (5 numbers each) and all the standard tools to properly run the system.</p>		
	<p>User training: Must provide user training (including how to use and maintain the equipment's) to the operator</p>		
	<p>Warranty :Warranty Period shall be 1 year</p>		
C	<p>Mould room equipments</p>		
	<p>Hot air/Steam bath oven for preparing all standard masks head, head and shoulder, thoracic and pelvis). Must have adjustable temperature/time with digital display.</p>	1 set	
	<p>Mould room alloy melter</p>	1 set	
	<p>Bolus sheet 30"x30"x0.5" cm</p>	3nos	
	<p>Bolus sheet 30"x30"x1" cm</p>	6nos	
	<p>Bolus sheet 30"x30"x1.5" cm</p>	4nos	
	<p> Wax bolus</p>	10pct	
	<p> Vacuum bag full body - pediatric</p>	2	
	<p> Vacuum Diaphragm pump system</p>	1	

	Storage Cart for Vacuum Bag	1	
	Tungsten Eye Shielding set (small, medium & large	2	
D	TSET		
	A complete set for patient support system: Frame/case, Auxiliary support, thin and wide fabric belts for chest and under arms, eye shields, etc	1 set	
E	Base plate (black) compatible to O-type mask .Must be a strong, lightweight Carbon Fiber and is CT compatible.	2 sets	
F	Breast board : have two arms cups, two wrist cups, two hand posts, a removable head cup, adjustable bottom-stopper, and a leg cushion	1 set	

