

# BP Koirala Memorial Cancer Hospital

## Bharatpur Chitwan

### Amendment Notice for Hospital Laundry Set Up and CSSD Autoclave Set UP IFB No. BPKMCH/NCB/40/081/082

**Amended on: March 21, 2025**

Points	Existing Clause	Amended/Added Clause
Bid Document Fee	Rs.3,000.00 ( In words Nepali Rupee Three Thousand Only )	<b>Rs.5,000.00 (In words Nepali Rupee Five Thousand Only )</b>
Technical specification of High Pressure vacuum steam sterilizer Point 4.9	Shall have fully automatic steam generator made of SUS304/AISI 304L chamber to feed steam to autoclave jacket and gasket groove. Water reservoir, water sensing electrodes, pressure switches and safety valve must be part of steam generation unit. It shall come with heating element of (20-25 KW).	<b>Shall have fully automatic steam generator made of SUS304/AISI 304L chamber to feed steam to autoclave jacket and gasket groove. Water reservoir, water sensing electrodes, pressure switches and safety valve must be part of steam generation unit. It shall come with heating element of 20 KW or more</b>
Technical specification of High Pressure vacuum steam sterilizer Point 8.3	Should use Float sensor for water level	<b>Should use Float sensor/Level Sensor for water level</b>
Technical specification of High Pressure vacuum steam sterilizer Point 8.5	Sterilization cycles: 20 Cycles	<b>Sterilization cycles: approx. 20 Cycles or more</b>
Technical specification of Washer Extractor 1.7 - 4	4. The basket volume should be of minimum 1000L , depth 790 mm along with diameter 1200 mm	<b>4. The basket volume should be of minimum 1000L , depth 745 mm or more along with diameter 1200 mm or more</b>

**Please visit the PPMO eGP system [www.bolpatra.gov.no/egp](http://www.bolpatra.gov.no/egp) and amended bid document for more details.**